

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: LJ380172G

Date Tested: 9-19-16

Instrument Number: 60543

Mac Address: 00:0A:8D:00:EC:7F

IP Address: 192.168.111.222

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	2.876	384	6.017	886	9.241	1384
Channel 2	2.795	383	6.016	883	9.244	1383
Channel 3	2.797	383	6.022	884	9.244	1383
Channel 4	2.793	383	6.018	883	9.245	1384
Channel 5	2.796	383	6.015	883	9.242	1383
Channel 6	2.801	383	6.019	884	9.243	1383

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 1897018

Calibration Date: 9-19-16

Precision digital multimeter SN: 36049893

Calibration Date: 9-19-16

ProHeat 35 power supply used for signal output SN: ME270375G

Calibration date: 9-19-16

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.

Technician: Jared Lamar

Calibration due date: 9-19-17

Signature