

Red-D-Arc Welderrentals

Calibration Certificate

Miller ProHeat 35

Unit Serial Number: LK230184G

Date Tested: 08/05/2013

Certified Calibration

Primary Standard						
	382° (F) 2.787VDC@RC9		882° (F) 6.013VDC@RC9		1382° (F) 9.239VCD@RC9	
	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display	RC9 output DC voltage reading	ProHeat 35 display
TC1	2.798	382	6.019	882	9.241	1382
TC2	2.787	382	6.010	882	9.237	1382
TC3	2.787	382	6.004	882	9.239	1382
TC4	2.788	382	6.009	882	9.237	1382
TC5	2.788	382	6.016	882	9.235	1381
TC6	2.782	382	6.011	882	9.237	1381
Allowable	2.749-2.825	+/- 3° F	5.975-6.051	+/- 3° F	9.201-9.277	+/- 3° F

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055

Calibration Date: 26/07/2012

Precision Digital Voltage Meter SN: 2142011

Calibration Date: 16/08/2012

Machine re-calibration due date: 08/05/2014

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician: Gino Rodin Date: 08/05/13

Signature

Montreal Canada office