

Red-D-Arc Welderrentals

Calibration Certificate

Unit Serial Number: MA120025G
 Instrument Number: 99291
 IP Address: 192.168.111.222

Date Tested: 06/16/15
 Mac Address: 00:0A:8D:01:83:DB

Primary Standard						
	382 (F)		882 (F)		1382 (F)	
	Signal Voltage	Recorder display	Signal Voltage	Recorder display	Signal Voltage	Recorder display
Channel 1	382	383	882	883	1382	1383
Channel 2	383	382	883	883	1382	1382
Channel 3	383	383	883	883	1383	1383
Channel 4	384	383	883	883	1383	1383
Channel 5	383	383	883	883	1383	1384
Channel 6	384	384	883	884	1383	1384

Allowable difference in temp reading is +or- 5 Deg. F.

Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2118055 Calibration Date: 24/07/14
 Precision digital multimeter SN: 2142011 Calibration Date: 24/07/14
 ProHeat 35 power supply used for signal output SN: MC450063G Calibration date: 06/16/15

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.
 Calibration due date: 06/16/16

Technician: G. We Rock
 Signature