

# Red-D-Arc Welderrentals

## Calibration Certificate

Unit Serial Number: MB430142G

Date Tested: 19-Jun-17

Instrument Number: 128510

Mac Address: 00:0A:8D:01:F5:FE

IP Address: 192.168.111.222

| Primary Standard |                |                  |                |                  |                |                  |
|------------------|----------------|------------------|----------------|------------------|----------------|------------------|
|                  | 382 (F)        |                  | 882 (F)        |                  | 1382 (F)       |                  |
|                  | Signal Voltage | Recorder display | Signal Voltage | Recorder display | Signal Voltage | Recorder display |
| Channel 1        | 2.789          | 382              | 6.019          | 883              | 9.236          | 1382             |
| Channel 2        | 2.788          | 382              | 6.008          | 881              | 9.235          | 1382             |
| Channel 3        | 2.791          | 383              | 6.017          | 881              | 9.244          | 1383             |
| Channel 4        | 2.791          | 381              | 6.019          | 883              | 9.246          | 1383             |
| Channel 5        | 2.788          | 382              | 6.020          | 883              | 9.243          | 1383             |
| Channel 6        | 2.793          | 383              | 6.024          | 883              | 9.250          | 1384             |

**Allowable difference in temp reading is +or- 5 Deg. F.**

### Instruments Used:

Fluke 714 Thermocouple Calibrator SN: 2097001

Calibration Date: 13-Mar-17

Precision digital multimeter SN: 2142012

Calibration Date: 13-Mar-17

ProHeat 35 power supply used for signal output SN: ME170131G


Calibration date: 19-Jun-17

Red-D-Arc does hereby certify the above instrument was calibrated against standards maintained by Red-D-Arc and meets or exceeds all published specifications. The accuracy of these standards is directly traceable to the National Institute of Standards and Technology.

Unit passed inspection.



Technician:



Calibration due date: 19-Jun-18

Signature